

Calibration certificate No: xxxxxx

Customer:
Manufacturer: Meatest
Type: M910
Serial number: xxxxxx
DN (diameter): 40 mm
Range [m³/h]: 13

Test conditions:

Test was done by a comparison with the standard flowmeters using volumetric method and flying start.

Standard flowmeters: YOKOGAWA AXF 015G DN15, YOKOGAWA AXF 015G DN80
Meter position: horizontal
Test medium: clean water
Water temperature in test: 21°C ± 1°C
Uncertainty of test: 0.3%
Output setting during test: 1000 Hz/Qmax

Measured values:

<u>Range in %</u>	<u>Deviation [%]</u>
50	-0.01
25	0.17
10	-0.13

Date: 04.02.2016

MEATEST, spol. s r.o.
K. Ambrožová